



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/748,791	12/26/2000	Par Lindh	279.352US1	3323
21186	7590	08/23/2005		
SCHWEGMAN, LUNDBERG, WOESSNER & KLUTH, P.A. P.O. BOX 2938 MINNEAPOLIS, MN 55402-0938			EXAMINER EVANISKO, GEORGE ROBERT	
			ART UNIT	PAPER NUMBER
			3762	

DATE MAILED: 08/23/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

09/748,791

Applicant(s)

LINDH ET AL.

Examiner

George R. Evanisko

Art Unit

3762

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 07 July 2005.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-6 and 13-20 is/are pending in the application.
- 4a) Of the above claim(s) 1-6, 13 and 14 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 15-20 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☐ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date _____
- ☐ Notice of Informal Patent Application (PTO-152)
- ☐ Other: _____

DETAILED ACTION

Election/Restrictions

Newly submitted claims 1-6, 13, and 14 are directed to an invention that is independent or distinct from the invention originally claimed for the following reasons: The inventions are related as combination (original claim 1) and subcombination (amended claim 1). Inventions in this relationship are distinct if it can be shown that (1) the combination as claimed does not require the particulars of the subcombination as claimed for patentability, and (2) that the subcombination has utility by itself or in other combinations (MPEP § 806.05(c)). In the instant case, the combination as claimed does not require the particulars of the subcombination as claimed because the combination does not require determining a duration of a PR interval. The subcombination has separate utility not requiring displaying the information, but as a method of directing a pacemaker to pace the heart. In addition, the inventions are related as process (claim 1) and apparatus (claim 15) for its practice. The inventions are distinct if it can be shown that either: (1) the process as claimed can be practiced by another materially different apparatus or by hand, or (2) the apparatus as claimed can be used to practice another and materially different process. (MPEP § 806.05(e)). In this case the process as claimed can be practiced by another and materially different apparatus not requiring a display screen to display the information or a programmer, but a control system for setting the pacing rate internally in a pacemaker.

Since applicant has received an action on the merits for the originally presented invention, this invention has been constructively elected by original presentation for prosecution on the merits. Accordingly, claims 1-6, 13, and 14 are withdrawn from consideration as being directed to a non-elected invention. See 37 CFR 1.142(b) and MPEP § 821.03.

Claim Rejections - 35 USC § 112

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

Claims 15-20 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. The subject matter which was not described in the original specification is the selection of an AV delay interval for delivering one or more ventricular pacing pulses with control circuitry using “a QRS complex”, intrinsic PR interval, “ and the timing relationship between right and left ventricular depolarizations”. The original specification only discussed using two different intrinsic PR intervals to select an AV delay. This rejection is related to new matter.

Claims 15-20 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the enablement requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to enable one skilled in the art to which it pertains, or with which it is most nearly connected, to make and/or use the invention. The subject matter which was not described in the specification is the control circuitry selecting an AV delay interval for delivering one or more ventricular pacing pulses using “a QRS complex, intrinsic PR interval, and the timing relationship between right and left ventricular depolarizations”, in combination with the other elements in the claim. Several questions arise in selecting the AV delay (as stated

Art Unit: 3762

in the claim), such as: How is the QRS complex used?; How is the timing relationship between right and left ventricular depolarizations used?; and how is the PR interval used alone and/or (and what is the relationship) with the QRS and ventricular depolarizations to produce the AV delay? This rejection is related to enablement.

Claims 15-20 are rejected under 35 U.S.C. 112, first paragraph, as based on a disclosure which is not enabling. Basing the AV delay on two different PR intervals (the most recent PR interval and the previously indicated PR interval) is critical or essential to the practice of the invention, but not included in the claim(s) is not enabled by the disclosure. See *In re Mayhew*, 527 F.2d 1229, 188 USPQ 356 (CCPA 1976). The specification describes on pages 14-16 the use of the most recent PR interval and the previously indicated PR interval to determine the AV delay, and therefore more than one PR interval is needed.

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 15-20 are rejected under 35 U.S.C. 112, second paragraph, as being incomplete for omitting essential elements, such omission amounting to a gap between the elements. See MPEP § 2172.01. The omitted elements are: the second PR interval used to determine the AV delay. As seen on pages 14-16 of the specification, both PR intervals are needed to determine the AV delay.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

Art Unit: 3762


A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to George R. Evanisko whose telephone number is 571 272 4945. The examiner can normally be reached on M-F 6:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on 571 272 4955. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

GRE
August 21, 2005


GEORGE R. EVANISKO
PRIMARY EXAMINER

8/21/5